

CERTIFICATE OF CONFORMITY



A D T



Equipment : Mobile Cart Computer
Brand Name : ONYX Healthcare
Test Model No. : VENUS-192ET-A1-1010
Multiple Listing : VENUS-192xxxxxxxxxx
("x" is 0 to 9, A to Z or blank for marketing purpose only)
Applicant : Onyx Healthcare Inc.
Test Report No. : CE150313D06

We, **Bureau Veritas Consumer Products Services (H.K.) Ltd., Taoyuan Branch**, declare that the equipment above has been tested in our facility and found compliance with the requirement limits of applicable standards, in accordance with the Directive 93/42/EEC. The test record, data evaluation and Equipment Under Test (EUT) configurations represented herein are true and accurate under the standards herein specified.

EN 60601-1-2:2007 +AC:2010

The EMC limits and test methods are referred to the following standards:

Emission:

CISPR 11:2009 +A1:2010 (Group I, Class B)

EN 61000-3-2:2014

EN 61000-3-3:2013

Immunity:

IEC 61000-4-2:2008 ED.2.0

IEC 61000-4-3:2010 ED.3.2

IEC 61000-4-4:2012 ED.3.0

IEC 61000-4-5:2005 ED.2.0

IEC 61000-4-6:2013 ED.4.0

IEC 61000-4-8:2009 ED.2.0

IEC 61000-4-11:2004 ED.2.0

NOTE: The above IEC basic standards are applied with latest version if customer has no special requirement.

Henry Lai / Director

Mar. 31, 2015

No. 47-2, 14th Ling, Chia Pau Vil., Lin Kou Dist., New Taipei City, Taiwan (R.O.C.)

Tel: 886-2-26052180 Fax: 886-2-26051924

<http://www.bureauveritas-adt.com> E-Mail: service.adt@tw.bureauveritas.com

