


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product	Mobile Medical Assistant Tablet
Name and address of the applicant	Onyx Healthcare Inc. 2ND FL, 135 Ln 235 PAO CHIAO RD, HSIN TIEN, New Taipei 231, Taiwan
Name and address of the manufacturer	Onyx Healthcare Inc. 2ND FL, 135 Ln 235 PAO CHIAO RD, HSIN TIEN, New Taipei 231, Taiwan
Name and address of the factory	Onyx Healthcare Inc. 2ND FL, 135 Ln 235 PAO CHIAO RD, HSIN TIEN, New Taipei 231, Taiwan. <input checked="" type="checkbox"/> Additional Information on page 2
Notes: When more than one factory, please report on page 2	
Ratings and principal characteristics	See Page 2
Trademark (if any)	ONYX Healthcare Inc. 
Type of Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	xxxONYX-MD116xxxxxxxxxx, See Page 2
Additional information (if necessary may also be reported on page 2)	N/A <input type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1, IEC 60601-1-6(ed.3), IEC 60601-1-6(ed.3);am1
As shown in the Test Report Ref. No. which forms part of this Certificate	170601504 issued on 2017-12-11

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-12-15

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-69018-UL

Model Details:

xxxONYX-MD116xxxxxxxx (Where x can be 0 to 9, a to z, A to Z, -, or blank)

Factories:

AAEON Technology Inc.
2nd FL 125 Lane 235, Baoqiao Rd Xindian District, New Taipei 231,
Taiwan.

Ratings:

For Switching Power supply:
(FSP Group Inc., model: FSP090M-RHA)
Input: 100-240 Vac, 50-60 Hz, 1.5-0.6 A
Output rating: 12 Vdc, 7.5 A

For Mobile Medical Assistant Tablet:

Input rating: 12 Vdc, 7.5 A

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-12-15

Signature:

Jan-Erik Storgaard