



---

# CERTIFICATE

---

**Issued Date: Jul. 25, 2017**  
**Report No. : 1770099R-ITCEP01V01**

This is to certify that the following designated product

**Product** : Mobile Medical Assistant Tablet  
**Trade name** : onyx  
**Model Number** : xxxONYX-MD116xxxxxxxxxx(Where "x" is 0~9, A~Z,  
"- " or blank)  
**Company Name** : ONYX Healthcare Inc.

This product, which has been issued the test report listed as above in DEKRA Laboratory, is based on a single evaluation of one sample and confirmed to comply with the requirements of the following EMC standard.

**EN 55032: 2012+AC: 2013 Class B**

**EN 61000-3-2: 2014**

**EN 61000-3-3: 2013**

**EN 55024: 2010+A1: 2015**

**IEC 61000-4-2 Ed. 2.0: 2008**

**IEC 61000-4-3 Ed. 3.2: 2010**

**IEC 61000-4-4 Ed. 3.0: 2012**

**IEC 61000-4-5 Ed. 3.0: 2014**

**IEC 61000-4-6 Ed. 4.0: 2013**

**IEC 61000-4-8 Ed. 2.0: 2009**

**IEC 61000-4-11 Ed. 2.0: 2004**

TEST LABORATORY

Vincent Lin / Director



**DEKRA Testing and Certification Co., Ltd.**

No. 5-22, Ruishukeng, Linkou Dist., New Taipei City 24451, Taiwan (R.O.C.)

TEL: +886-2-8601-3788 FAX: +886-2-8601-3789 Email: [info.tw@dekra.com](mailto:info.tw@dekra.com) <http://www.dekra.com.tw>