





Test Report issued under the responsibility of:



Prodigy Technology Consultant Co., Ltd.

<b>IEC/EN 60601-1</b>	
<b>Medical electrical equipment</b>	
<b>Part 1: General requirements for basic safety and essential performance</b>	
<b>Report Reference No.</b> ..... :	M150408-08-A0
<b>Date of issue</b> .....	2015-06-11 .....
<b>Total number of pages</b> .....	197 .....
<b>Testing Laboratory</b> .....	Prodigy Technology Consultant Co., Ltd.
<b>Address</b> .....	No.181, Sec. 2, Wunhua 1st Rd., Linkou District, New Taipei City 244, Taiwan CHINESE TAIPEI
<b>Applicant's name</b> ..... :	ONYX HEALTHCARE INC
<b>Address</b> .....	2F., No.135, Ln. 235, Baoqiao Rd., Xindian Dist., New Taipei City 231, Taiwan
<b>Test specification:</b>	
<b>Standard</b> .....	EN 60601-1:2006/A12:2014; IEC 60601-1: 2005 + CORR. 1 (2006) + CORR. 2 (2007)
<b>Test procedure</b> ..... :	CB Scheme / Service in CE marking
<b>Non-standard test method</b> .....:	None
<b>Test Report Form No.</b> ..... :	DTL-091-A5
<b>Test Report Form Originator</b> .....	Prodigy Technology Consultant Co., Ltd.
<b>Master TRF</b> .....	Dated 2015-04-13
<b>Test item description</b> .....	Healthcare Infotainment Computer
<b>Trade Mark</b> ..... :	 or ONYX Healthcare Inc.
<b>Manufacturer</b> .....	Same as
<b>Model/Type reference</b> ..... :	ONYX-BE182xxxxxxxxxx (the "x" can be 0 to 9, a to z, A to Z, "-", or blank for marketing purpose only, and no impact safety related constructions and critical components.)
<b>Ratings</b> ..... :	1) 12Vdc, 5A 2) 24Vdc, 2.91A, 70 W max



<b>Testing procedure and testing location:</b>	
<input checked="" type="checkbox"/> <b>Testing Laboratory:</b>	
Testing location/ address .....	Prodigy Technology Consultant Co., Ltd. / No.181, Sec. 2, Wunhua 1st Rd., Linkou District, New Taipei City 244, Taiwan CHINESE TAIPEI
<input type="checkbox"/> <b>Associated Test Laboratory:</b>	
Testing location/ address .....	
Tested by (name + signature).....	Yama Cheng 
Approved by (+ signature) .....	Tony Hsu 